

Verification No.: CMSD21070532004



Verification

of Conformity with European Directives

Applicant: Shandong Chenhong Medical Technology Co., Ltd.
Address of Applicant: East of Nanshou Road, Jinlan Road, Qufu City, Jining City, Shandong Province, China
Product Name: Electric Operating Table
Model: CH-T100, CH-T200, CH-T300, CH-T400, CH-T500, CH-T600
Test Requested: Medical Device Regulation(EU) 2017/745(MDR)
 EN 1041:2008+A1:2013, EN ISO 10993-1:2020, EN ISO 10993-5:2009,
 EN ISO 10993-10:2013, EN ISO 15223-1:2016, EN ISO 14971:2019,
 EN 60601-1:2006+A12:2014, EN IEC 60601-2-46:2019

Classification
of medical device: Medical Devices – Class I
TCF Reference: TMSD21070532004
Date of Issue: July 09th, 2021
Date of expiry: July 08th, 2026

The submitted sample of the above product has been tested for CE marking according to the following European Directive(s):

Medical Device Regulation(EU) 2017/745(MDR)

Conclusion

The referred test report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements.

The CE marking as shown below can be affixed on the product after preparation of necessary technical documentation.

Approved by: Hermann Weiher

For and on behalf of
Global Testing Services Co., Ltd



Global Testing Services Co., Ltd
 E-mail: info@gts-lab.com <http://www.gts-lab.com>
 Floor 2nd, Building D-1, No. 128, Shenfu Road, Minhang District, Shanghai, China.

Copyright of this certificate is owned by Global Testing Services Co., Ltd and may not be reproduced other than in full and with the prior approval of the General Manager.